

THE NEW FACE of Glucometer®

Ascensia ELITE™

Diabetes Care System

USER GUIDE

For Use With Ascensia ELITE™ Blood Glucose Meter

Need Help With Reimbursement? Call Our Bayer Care Help Line: 1-877-787-1912.
If You Have Any Questions, Comments or Need Help With Reimbursement For
Your Testing Supplies, Please Call 1-800-348-8100.



Bayer 

Introduction

What it is:

The Ascensia *ELITE*[™] Diabetes Care System is a blood glucose monitoring system. The system is made up of the Ascensia *ELITE*[™] Meter, Test Strip, and Code Strip. For quality control of the system, the Ascensia *ELITE*[™] Check Strip and Normal, Low and High Controls are also available.


How it works:

The Ascensia *ELITE* blood glucose testing process is based on an electrode sensor technology. Capillary action at the end of the Test Strip draws a small amount of blood into the reaction chamber and a reading is displayed in 30 seconds. No timing, wiping or blotting is required.

What is new and different:

Ascensia *ELITE*[™] Test Strips are calibrated to give plasma/serum equivalent glucose results. Plasma/serum results are used by doctors and laboratories as the most accurate method for reading blood glucose levels. The Ascensia *ELITE* System provides you with readings that correspond directly to those of your doctor or clinic. Compare Ascensia *ELITE* Test Strip results only to systems which use a plasma/serum calibration.

TESTING MATERIALS




-
- 1 **Test Slot** – provides an opening for inserting strip into the Meter.
 - 2 **Icon** – shows proper orientation and direction for inserting the Test Strip, Code Strip or Check Strip into the Meter.
 - 3 **Display** – displays test results, current Code (Function) Number (F- #), or low battery indicator.
 - 4 **Batteries** – provide power for the Meter.
 - 5 **Check Strip** – confirms that the Meter electronics are working properly, and activates the memory recall and the delete option.
 - 6 **Code Strip** – codes (calibrates) the Meter.
 - 7 **Test Strip** – tests for glucose in blood and control solutions, and can be used to “mark” a control test or to delete a test result from the memory.
 - 8 **Test end** – draws sample into the reaction chamber of the Test Strip.
 - 9 **Meter end** – inserted into the Meter to activate power.
 - 10 **MICROLET® Adjustable Lancing Device** and 11 **lancet** – for skin puncture.
 - 12 **Normal Control Solution** – a solution with a known amount of glucose used to confirm that the system is working properly (Low and High Control Solutions are also available).

Introduction (continued)

What you do:

1. Insert the **Test Strip** into the **Test Slot** located at the top, center of the Meter.
2. Touch the **test end** (tip) of the **Test Strip** to a drop of blood. A “beep” sounds after the **Test Strip** reaction chamber is filled with blood. The test result is displayed in 30 seconds.



The Meter turns on when the Test Strip is inserted and turns off when the Test Strip is removed, therefore, no operating buttons are required. The Meter automatically stores the last 20 test results, including both blood glucose and control test results.

The Ascensia *ELITE* System provides blood glucose results ranging from 20–600 mg/dL (1.1–33.3 mmol/L). The Meter is battery operated and approximately 1000 tests may be performed before replacing batteries.


CAUTION: Before using any product to test your blood glucose (sugar), read all instructions and practice the test. Do all quality control checks as directed and consult with a diabetes health professional. These recommendations apply to all blood glucose monitoring systems and are supported by the American Association of Diabetes Educators,* the American Diabetes Association, the Food and Drug Administration and the Health Industry Manufacturers Association.

*You may call the American Association of Diabetes Educators for the location of a Certified Diabetes Educator (CDE) near you 1-800-338-DMED.

Special Feature


1. WHAT IS DIFFERENT?

The part of the Test Strip that touches the sample.




2. HOW IS IT DIFFERENT?


The blood sample is drawn *into* the Test Strip and *is not* put on the Test Strip.



3. WHEN DOES THIS HAPPEN?





The test end (tip) of the Test Strip draws (pulls) the blood sample into the Test Strip *when* the tip touches the blood sample.





4. IN OTHER WORDS:

Think of the Test Strip as a straw used for drinking. You put the straw into the liquid. The liquid is drawn into the straw at one end...**SO** the blood sample is drawn into the Test Strip at one end when you put the tip of the Test Strip into the sample. We call this "Sip-in Sampling."



Contents

SECTION 1 – The Ascensia *ELITE*[™] Diabetes Care System

- Learn About the System 1
- Install/Replace Batteries 2
- Check the Meter Performance 3
- The Check Strip Test 4

SECTION 2 – Prepare for the Blood Glucose Test

- Code (Calibrate) the Meter 6
- The Control Test 9

SECTION 3 – The Blood Glucose Test

- Get Ready for the Blood Glucose Test 14
- Perform the Blood Glucose Test 17

Contents (continued)

SECTION 4 – Learn About the Memory Feature

- Introduction to Memory 21
- Recall Test Results from the Memory 23
- Delete **All** Test Results from the Memory 24
- Delete a **Single** Test Result from the Memory 25
- Mark a Control Test Result 26
- Things to Remember 27

SECTION 5 – Care of the Meter and Strips 29

SECTION 6 – Solve Problems

- Meter and Test Strip Problems 31
- Test Result Problems 33

SECTION 7 – Service the System

- Service 35
- Check List 36
- Limitations of the Procedure 37
- Specifications 38
- Supplies 40
- Warranty 41

Learn About the System

SECTION 1

This booklet is designed to guide you through the operation of the Ascensia *ELITE*™ Diabetes Care System with detailed directions and illustrations. A Quick Reference Guide is provided, and can be carried in your Meter Carrying Case.

Please read the entire User Guide before you begin testing.

Before reading the User Guide, you will need to unpack and set up your testing materials. See the “Testing Materials” foldout, attached to the first page of this User Guide, to identify testing materials needed.



Install / Replace Batteries

SECTION 1

THE BATTERIES: The Meter uses two (3-volt lithium) batteries. These are furnished with the Meter.


WHEN TO REPLACE: When the battery voltage becomes weak (after approximately 1000 readings), the Meter will show a battery display (■). When this display appears, it is time to replace the batteries (see page 40 for battery types). The Meter will continue to provide accurate results for a period of time until the batteries are exhausted. At this time, the Meter will “lock you out” from further testing.

DIRECTIONS:

1. With the tip of a finger, press the battery compartment lid latch toward the word “open.” Lift the lid.
2. Place the two new batteries into the battery compartment with the “+” side up and close the lid. **NOTE:** If replacing batteries, remove the old batteries from the compartment and discard.

Things to remember:

- **WARNING:** Keep batteries away from children. Lithium batteries are poisonous. If swallowed, immediately contact your physician.
- Discard batteries according to your local environmental regulations.
- The Meter memory will not be erased when batteries are changed.



Check the Meter Performance

WHEN:

When your Meter is new and anytime you want to check that your Meter is functioning properly.

WHAT TO EXPECT:

A Check Strip test result that falls within the Check Strip Range tells you that the Meter electronics are working properly.

Note the **Check Strip Range** printed on the check strip package insert.

A Check Strip test result that falls outside the Check Strip Range tells you the Meter may not be functioning properly. If this occurs, refer to Section 6 – page 31.

OTHER:

The Check Strip can also be used to activate the Meter's memory recall function, or memory deletion option (see Section 4).




The Check Strip Test


MATERIALS REQUIRED:

- Ascensia *ELITE*™ Meter
- Ascensia *ELITE*™ Check Strip


CHECK THE METER:



1 Remove the Check Strip from its package.




2 Insert the Check Strip fully into the Meter as shown (with tab pointing right). A beep sounds and a full display* appears. Another beep sounds and the Check Strip test result is displayed. A check mark (✓) shows on the display when the Check Strip test is done.



*A full display, as shown here, indicates that all characters in the display are working properly. If ever all the display segments *do not* appear, see Section 6, page 32.

The Check Strip Test (continued)



Things to remember:

- It is important to insert the Check Strip with a firm motion until it comes to a full stop. When inserted fully, the tab on the top portion of the strip should be completely hidden within the Meter.
- Always keep the Check Strip stored in its package when not in use.
- Do not bend the Check Strip.
- You can also use the Check Strip to recall the Meter's memory, or to delete all of the stored test results from the memory (see Section 4).

3 If the Check Strip test result is within the range listed on the package insert found inside the Check Strip package, the Meter is functioning properly. If not, refer to the Problem Solving Chart in Section 6.


Carefully remove the Check Strip from the Meter, and replace it in its package.

Code (Calibrate) the Meter

WHY CODE THE METER?

To match the Meter to the reactivity of the Test Strip, you must first code (calibrate) the Meter using the Code Strip provided in the carton of Ascensia *ELITE*[™] Test Strips now in use.

Each lot of Test Strips is tested extensively and assigned a Function Number (0 through 12) identified on the Code Strip.



WHEN TO CODE THE METER


Reset the code for each new carton of Ascensia *ELITE*[™] Blood Glucose Test Strips used, unless the new carton of Test Strips has the same Function Number as the previous one.

Code the Meter (continued)


MATERIALS REQUIRED:

- Ascensia *ELITE*™ Meter
- Ascensia *ELITE*™ Code Strip

CODE THE METER:




1 Open your new carton of Test Strips and find the clear plastic packet containing the Code Strip.



2 Carefully tear open the packet, and remove the Code Strip. Save the packet to store the Code Strip.


Code the Meter (continued)



3 Insert the Code Strip into the Test Slot (as shown). A beep sounds and a full display appears briefly. Another beep sounds and the Function Number (for example, **F-4**) appears in the Display.

Things to remember:

- It is important to insert the Code Strip with a firm motion until it comes to a full stop. When inserted fully, the tab on the top portion of the strip should be completely hidden within the Meter.
- The Function Number appearing in the Display **must** match the number appearing:
 - on the Code Strip.
 - on the front of the Code Strip packet (see #2).
 - on the back of the Test Strip packet.
- If the Function Numbers match as indicated above, the Meter may now be used to run a control test or blood glucose test.
- The Meter must be coded again when a new carton of Test Strips is used unless the F# of the new carton is the same as the old carton. Use the new Code Strip found in the new carton of strips.
- Store the Code Strip in the clear plastic packet until all the strips are used (in case you ever need to recode the meter). Throw away the old Code Strip when Test Strips are all used.



To make sure your Ascensia ELITE™ Diabetes Care System is performing properly, a Control Test is performed using the Ascensia ELITE™ Normal Control solution.

WHY? A Control Test confirms that the system is functioning properly and you are using proper technique. When running a Control Test, the system is defined as: the Meter and the Test Strips. In addition to the Normal Control, an Ascensia *ELITE* Low and High Control are also available for use with this system.

WHEN TO RUN A CONTROL TEST:

- Any time you want to confirm that the Meter system is working properly, and that your testing technique is correct.
- Any time you question your blood glucose test results.
- On the advice of your physician or healthcare professional.

WHAT TO EXPECT:

- When the Control Test is done properly, the Control Test result should fall within the range listed on the end flap of your Test Strip Carton.
- A Control Test result that falls outside of the specified range means the system may not be functioning properly or your testing technique may not be correct. If this occurs, refer to Section 6 of this User Guide or the Ascensia *ELITE* Control package insert for assistance.

MARK THE CONTROL TEST RESULT (if using the memory feature):

You should always “mark” a Control Test result with a “C” as described in the following procedure. Marking with a “C” keeps a Control Test result from being included in the average of your blood glucose tests.


The Control Test (continued)

MATERIALS REQUIRED:

- Ascensia *ELITE*™ Meter
- Ascensia *ELITE*™ Normal Control Solution
- Ascensia *ELITE*™ Test Strip and Code Strip
- CLINILOG® Diary

PREPARE FOR TEST:

If not already done, code your Meter using the Code Strip (see page 7). Also check to make sure the Test Strips and the Normal Control Solution have not passed their expiration date.




1 Remove foil packets from the Test Strip carton and tear off a single packet.



2 To open the Test Strip packet, carefully PEEL the foil until the Test Strip is completely exposed (as shown).


NOTE: The numbers printed on each Test Strip are manufacturing numbers and should not be confused with a Function Number.

The Control Test (continued)




3 Remove the Test Strip from packet. Holding the round end, insert the strip fully into the Meter.

NOTE: Save the empty foil packet for use in Step 6.



4 A beep sounds and a full display appears, followed by the Function Number (F#) and the previous test result. The Function Number and the previous test result begin flashing alternately.


If you are not using the memory feature, go directly to Step 6.




5 Mark your Control Test result by pulling out the Test Strip, and *quickly* inserting it back into the Meter (within 2 seconds). A C- and the Function Number F# begin flashing in the Display. Marking a Control Test excludes the Control Test result from the average of stored Blood Glucose Test results.

The Control Test (continued)


SECTION 2



6 Gently squeeze a small drop of Control solution onto the inside of the foil you saved in Step 3.




7 Touch and hold the TEST END of the Test Strip in the drop until after the Meter beeps. A small amount of Control solution is automatically drawn into the Test Strip. The timer begins counting down from 29 seconds.



8 After 29 seconds, the Control Test result appears in the Display. Compare the result to the range listed on the end flap of your Test Strip carton. Using a tissue, remove the Test Strip and discard. The Meter will turn off automatically.

The Control Test (continued)



FOLLOW UP:

Record and date all Control Test results in your CLINILOG Diary.

CAUTION: If the Control Test result does not fall within the stated range, the system may not be operating properly. See Section 6 of this User Guide and the Ascensia *ELITE*[™] Control package insert for help.

Things to remember:

- Run a Control Test frequently to make sure that you are using good technique and that your system is working properly.
- Always check the EXPIRATION DATE on the Test Strip foil packet or carton before using. Throw away expired Test Strips. Use only unexpired Test Strips.
- Use the Test Strip immediately after removing from the foil packet.
- Insert the Test Strip until it comes to a full stop. When inserted fully, the tab on the top portion of the strip should be completely hidden within the Meter.
- If using the memory feature, be sure to “mark” the Control Test *before* running the test (see Section 4).
- Remove the Test Strip immediately after use to turn off the Meter and preserve battery life.
- The Meter turns off automatically after 3 minutes if the strip is left in the Meter.
- Test Strips cannot be reused.
- The end flap of your Test Strip carton lists *specific* Control Ranges for that carton of strips. Please keep this end flap until all the Test Strips in the carton are used.




Get Ready for the Blood Glucose Test

MATERIALS REQUIRED:

- Ascensia *ELITE*™ Meter
- Ascensia *ELITE*™ Test Strip and Code Strip
- MICROLET® Adjustable Lancing Device/Lancet
 - Tissue or cotton ball for wiping finger after testing
- CLINILOG® Diary

PREPARE FOR TEST:


If not already done, code your Meter using the Code Strip (see page 7). Also check to make sure the Test Strips have not passed their expiration date.




- 1** Load the MICROLET Adjustable Lancing Device:
See the MICROLET Adjustable Lancing Device package insert for complete instructions. Place the MICROLET Adjustable Lancing Device on a clean surface while preparing for your blood glucose test.

Get Ready for the Blood Glucose Test (continued)

SECTION 3



2 Wash your hands. Use warm soapy water. Rinse and dry thoroughly.



3 Remove foil packets from carton and tear off single packet.




4 To open the Test Strip packet, carefully peel the foil until the Test Strip is completely exposed (as shown).

NOTE: The numbers printed on each Test Strip are manufacturing numbers and should not be confused with a Function Number.


Get Ready for the Blood Glucose Test (continued)

SECTION 3



5 Remove Test Strip from packet. Holding the round end, insert the strip fully into the Meter.

NOTE: Be sure to save the foil to use when disposing of the Test Strip.




6 A beep sounds and a full display appears, followed by the Function Number (F#) and the previous test result. The Function Number and the previous test result begin flashing alternately.

Proceed immediately (within 3 minutes) to Step 1 on the next page...

Perform the Blood Glucose Test

SECTION 3



1 Stick your finger with the MICROLET Adjustable Lancing Device and form a small drop of blood.


Note: Multiple Site Testing

Consult with your healthcare professional prior to testing from a site other than your fingertips.


Blood glucose testing results from sites other than your fingertip could be significantly different. These differences may vary from individual to individual, and are physiological differences caused by factors such as food intake, health state and medication being used.

If you perform testing other than from a fingertip:

- Select a soft, fleshy area of skin that is free from hair and visible veins.
- Wash the puncture site with soap and warm water. Rinse and dry thoroughly. This ensures cleanliness and increases blood flow to the puncture site.
- After you have determined the proper depth of your selected site, set the lancing device for that depth of puncture. Firmly press the lancing device against the skin at the selected lancing site, then press the button to lance the skin. When a drop of blood forms on the surface of your skin, bring the meter and test strip up to touch the drop.





Perform the Blood Glucose Test (continued)




2 Touch and hold the test end (tip) of the Test Strip to the drop of blood until after the Meter beeps. See page iv for the right method.

Blood is automatically drawn into the Test Strip. The timer begins counting down from 29 seconds.



3 After 29 seconds, your blood glucose result appears in the display. The test result remains in the display.


NOTE: If **Lo** appears in the display, your blood glucose value may be below 20 mg/dL (1.1 mmol/L). If **Hi** appears, your result may be above 600 mg/dL (33.3 mmol/L). See page 33 for other reasons **Lo** or **Hi** could occur. *Follow the advice of your medical team for hypoglycemia (low blood sugar) if your blood glucose is below 50 mg/dL (2.8 mmol/L) or hyperglycemia (high blood sugar) if your blood glucose is above 250 mg/dL (13.9 mmol/L).*



4 If you want to store the test result, remove the Test Strip from the Meter as shown. The Meter turns off automatically and the test result is now stored in the memory. You may wish to use a tissue or the foil wrapper from the strip to remove the used strip.


CAUTION: Dispose of the used Test Strip and lancet carefully to prevent injury or contamination to others. You may use the empty foil packet to dispose of the Test Strip.

Perform the Blood Glucose Test (continued)



5 If you do not want to store the test result in the memory, pull out the Test Strip and *quickly* insert it back into the Meter (within 2 seconds).

dEL appears in the Display, and the test result will **not be stored** in the memory. Do not pull the Test Strip out of the Meter until the Display goes blank, or the test result *will not* be deleted.



FOLLOW UP:

Record the result in your CLINILOG Diary.

NOTE:

If you do not remove the Test Strip, the test result remains in the Display for 3 minutes.

Things to remember:

- Never use a Test Strip beyond the printed expiration date (on the foil packet).
- Use the Test Strip immediately after removing from the foil packet.
- It is important to insert the Test Strip until it comes to a full stop. When inserted fully, the tab on the top portion of the strip should be completely hidden within the Meter.
- After inserting Test Strip, wait until the F# and previous test result begin flashing before applying sample to strip.
- Always make sure the Function Number printed on the Code Strip, Code Strip Packet and the Test Strip Foil matches the Function Number displayed in the Display.
- To be sure your Test Strip fills completely, hold the test end of the Test Strip in the blood drop until after the “beep” sounds.
- Remove the Test Strip immediately after use to turn off the Meter and preserve battery life. The Meter turns off automatically after 3 minutes if the strip is left in the Meter.
- You can recall stored test results from the Meter memory by inserting a new Test Strip (before applying blood), and waiting 3 minutes. The memory recall function will then begin automatically. Do not recall results with a *used* Test Strip.
- Teach another person to use the system so you will have someone to help should you ever need assistance.
- If a problem occurs, refer to Section 6 or the Test Strip package insert.

This section will instruct you on the use of the Ascensia *ELITE*™ Meter's memory feature. The following questions and answers should help you learn about this feature.

How many test results does the memory store?

The memory stores up to 20 test results, and also provides the “average” of the Blood Glucose Test results currently stored in the memory. Both Blood Glucose Test results and Control Test results can be stored in the memory. If 20 test results are stored and a new one is added, the oldest test result is automatically deleted. The newest one stored will now be #1 in memory recall.

How do I recall the stored test results?

By inserting an Ascensia *ELITE*™ Check Strip into the Meter and waiting 10 seconds. After the 10 second period, an **A-** appears in the display followed by the average of all the stored test results. If a Check Strip is not available, you may use a new Test Strip to access the memory. When you use a Test Strip, you must wait 3 minutes after inserting the Test Strip to access the memory.

Can I delete test results?

Yes, you can delete either all of the test results stored in the memory, or a single test result before it goes into the memory.


How will I know the difference between a stored Blood Glucose Test and a Control Test?

A Control Test result in memory is preceded by a C- display.

Since Control Test results should not be a part of the average blood glucose result, marking the Control Test result with a C- tells the Meter not to use that result when calculating the average. The C- is entered at the beginning of the Control Test.


The following pages will instruct you on performing each one of these features.

Recall Test Results from the Memory




1 Insert the Check Strip fully into the Meter as shown.


NOTE: If your Check Strip is not available, you can use a *new* Test Strip for this procedure. However, you have to wait 3 minutes for the memory sequence to begin.



2 A beep sounds and a full display appears briefly. Another beep sounds and the Check Strip result is displayed.



3 After 10 seconds, the memory sequence begins automatically. An **A-** (for average) appears for 2 seconds, followed by the average test result value (for 4 seconds).



4 The memory recall sequence continues. The most recent result is preceded by a **1-** followed by all other stored results. The last result displayed is the oldest test result.

NOTE: If no values are stored, three bars --- appear after the average **A-**.


Delete ALL Test Results from the Memory

SECTION 4




1 Insert the Check Strip fully into the Meter. Wait 10 seconds for the memory sequence to begin.

NOTE: You may also use a Test Strip to delete all test results from memory. However, you must wait 3 minutes for the memory sequence to begin. When you see the memory sequence begin, proceed to Steps 2 and 3.



2 Pull out the Check Strip, and *quickly* insert it back into the Meter (within 2 seconds).




3 The Display begins flashing **dEL** and **ALL** every second for 5 times, and then the Display goes blank.

Do not pull the Check Strip out of the Meter until the Display goes blank, or the stored test results will not be deleted.


Delete a SINGLE Test Result from the Memory

SECTION 4




1 A test result may be deleted at the end of a test during the 3 minutes it is displayed. Use a tissue to pull out the Test Strip and quickly insert it back into the Meter (within 2 seconds).

NOTE: Single test results can only be deleted at the end of a test.



2 dEL appears in the Display for 2 seconds and then the Meter turns off.


NOTE: Do not pull the Test Strip out of the Meter until the Display goes blank, or the test result will not be deleted.



3 After the Meter turns off, remove and discard the used Test Strip.


Mark a Control Test Result

SECTION 4




1 To prevent the Control results from being included in the average, you should mark each Control Test with a "C." This must be done before performing the Control Test.

First, insert the Test Strip fully into the Meter.



2 A beep sounds and a full display appears briefly. The Function Number and the **previous test result** begin flashing alternately.



3 Pull out the Test Strip, and *immediately* insert it back into the Meter (within 2 seconds). A C- and the F# begin flashing alternately in the Display.

You can now run the Control Test (see page 9).

Recall Test Results from the Memory (see page 23)

- Low (below 20 mg/dL [1.1 mmol/L]) and high (above 600 mg/dL [33.3 mmol/L]) test results are displayed and stored in memory as **Lo** and **Hi**. In calculating the average, **Lo** is entered as 20 mg/dL (1.1 mmol/L), and **Hi** is entered as 600 mg/dL (33.3 mmol/L).
- A **C-** display preceding a test result display indicates that the test result is a Control Test result.

Delete Test Results from the Memory (see pages 24 and 25)

- The only time you can delete a single test result (i.e., keep it from being stored in the memory) is at the end of a Control or Blood Glucose Test.
- A test result is automatically stored in the memory if it is not deleted at the end of the test.
- If you decide not to delete the test result after beginning the delete process, pull out the strip while **dEL** is still displayed.

Mark a Control Test Result (see page 26)

1. You should “mark” your Control Test results with a “C” before storing them in the memory. This prevents the Meter from averaging your Control Test results in with your blood glucose results.
2. Each Control Test result stored in memory will take up one of the 20 memory spaces that are available. Therefore, if you have 3 Control Test results and 17 blood glucose test results stored in the current memory, the memory sequence might look like this:

E X A M P L E	}	A - IIIII▶ 139	5 - IIIII▶ 200	10 - IIIII▶ 103	16 - IIIII▶ 89
		1 - IIIII▶ 116	6 - IIIII▶ 189	11 - IIIII▶ 153	17 - IIIII▶ 92
		2 - IIIII▶ 154	7 - IIIII▶ 90	12 - IIIII▶ 195	C - IIIII▶ 99
		3 - IIIII▶ 102	C - IIIII▶ 102	13 - IIIII▶ 94	
		C - IIIII▶ 95	8 - IIIII▶ 184	14 - IIIII▶ 207	
		4 - IIIII▶ 92	9 - IIIII▶ 115	15 - IIIII▶ 190	

When the memory is full, new test results are entered into position #1 and the oldest result disappears from the memory.

3. When the Meter calculates the average, the Control Test results in the memory will not be a part of the average. In the above example, the average is figured using only the 17 blood glucose test results listed.

Care of the Meter and Strips

SECTION 5



1. To help keep the Meter and Test Strips free of oils and other contaminants, wash hands and dry thoroughly before using.



2. Handle the Meter carefully. Dropping may damage the electronics or cause other malfunctions.



3. Do not expose the Meter, Test Strip, Code Strip or Check Strip to excessive humidity, heat, cold, dust or dirt.



4. If the Meter exterior needs cleaning, dampen a facial tissue with fresh water and wipe the Meter and Display carefully. Avoid spilling any liquids on the Meter. Liquids may damage the Meter's electronics.



5. Store Meter in the case provided or a replacement case of your choosing whenever possible.




WARNING: Potential Biohazard: Healthcare professionals or persons using this system on multiple patients should be aware of the following and should follow the infection control procedure approved by their facility.

All products or objects which come in contact with human blood, even after cleaning, should be handled as if capable of transmitting viral diseases.

The user should follow the recommendations for prevention of blood-borne transmissible diseases in healthcare settings, as recommended for potentially infectious human blood specimens in National Committee for Clinical Laboratory Standards, *Protection of Laboratory Workers from Instrument Biohazards and Infectious Disease Transmitted by Blood, Body Fluids and Tissue: Approved Guideline*. NCCLS document M29-A [ISBN 1-56238-339-6] NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898, USA, 1997. This document has complete information on the area of user protection and can be used as background material for instruction.

Meter and Test Strip Problems

PROBLEM	CAUSE	SOLUTION
E-1 } E-2 } display E-4 }	Meter electronics failure.	Immediately, contact the Customer Service Dept. 1-800-348-8100. <i>The meter will no longer function.</i>
Display goes blank after Test, Code or Check Strip is inserted into the Meter.	<ol style="list-style-type: none">1. Batteries are not installed correctly.2. Strip is not fully inserted into the Meter.3. Batteries are dead.4. Too much time has passed since inserting the strip and the Meter has turned off.	<ol style="list-style-type: none">1. Install batteries correctly.2. Insert Strip until tab is inside Meter.3. Replace with new batteries.4. Remove and reinsert the strip.
The Function Number displayed is different from the number printed on the Test Strip foil packet or carton.	<ol style="list-style-type: none">1. Wrong Code Strip used.2. Code Strip not inserted properly.	<ol style="list-style-type: none">1. Use correct Code Strip. Discard incorrect one.2. Reinsert Code Strip carefully, but firmly.
Check Strip test result out-of-range.	<ol style="list-style-type: none">1. Check Strip (Meter end) dirty or damaged.2. Meter electronics failure.	<ol style="list-style-type: none">1. Wipe Check Strip thoroughly with clean dry tissue, and retest.2. Contact Customer Service Department.


PROBLEM	CAUSE	SOLUTION
Meter fails to start countdown.	<ol style="list-style-type: none">1. Blood applied wrong; see pp. iii, iv.2. Test Strip inserted incorrectly.3. Sample is applied while full display is showing .	<ol style="list-style-type: none">1. Form a larger drop of blood or control solution and rerun test using a new Test Strip. Hold test end to sample until after the “beep.”2. Insert Test Strip correctly (Meter end first). 3. Wait for F# and last test result to begin flashing before applying sample to strip.
Some display segments do not appear.	Meter electronics or battery failure.	Replace batteries and if Display is still incorrect, contact Customer Service Department.
 Battery display.	<ol style="list-style-type: none">1. Battery voltage low.	<ol style="list-style-type: none">1. Replace batteries. See Section 1, page 2.

PROBLEM	CAUSE	SOLUTION
<p>"Lo" display.</p> <p>CAUTION: Glucose levels below 50 mg/dL (2.8 mmol/L) may indicate a potentially serious medical condition.</p>	<ol style="list-style-type: none">1. The blood glucose value may be below 20 mg/dL (1.1 mmol/L).2. Test Strip not filled completely.3. Test Strip is defective or has deteriorated.4. Wrong Function Number used.	<ol style="list-style-type: none">1. Possible hypoglycemic condition (low blood sugar). Contact your physician or diabetes counselor about your condition.2. Rerun test with new strip and apply sufficient sample to fill Test Strip. Hold test end to sample until after the "beep."3. Rerun test using a new, unexpired Test Strip.4. Use correct Code Strip provided with carton of Test Strips being used.
<p>"Hi" display.</p> <p>CAUTION: Glucose levels above 250 mg/dL (13.9 mmol/L) may indicate a potentially serious medical condition.</p>	<ol style="list-style-type: none">1. The blood glucose value may be above 600 mg/dL (33.3 mmol/L).2. Same as #3 above.3. Same as #4 above.	<ol style="list-style-type: none">1. Possible hyperglycemic condition (high blood sugar). Contact your physician or diabetes counselor about your condition.2. Same as #3 above.3. Same as #4 above.

PROBLEM	CAUSE	SOLUTION
Blood glucose or control test results are inconsistent, or control test results are not within the specified range.	<ol style="list-style-type: none">1. Not enough blood or control solution is drawn into the Test Strip.2. Test Strips or control solution have passed their expiration date.3. Deteriorated Test Strip caused by heat or humidity.4. Meter, Control Solution and/or Test Strip are not at room temperature when used.	<ol style="list-style-type: none">1. Rerun test with new Test Strip and apply sufficient sample to fill Test Strip. Hold TEST END to sample until after the “beep.”2. Obtain new Test Strips and/or Control Solution that are within their expiration date.3. Run a Control Test using a new Test Strip. If results are still out of range, replace carton of Test Strips. See carton for proper storage instructions.4. Allow time for Meter, Control Solution and/or Test Strips to come to room temperature before use.

IMPORTANT: To have accurate test results, sufficient sample to fill the reaction chamber is needed. The following drawings show how a Test Strip looks when filled and not filled.

RIGHT
(filled)



WRONG
(not filled)



If attempts to correct a problem fail, please call our Customer Service Department. We have trained specialists to assist you.



**In the U.S.A., call toll free at:
1-800-348-8100**

IMPORTANT:

- Speak to a Customer Service Representative before returning your Meter for any reason. He/she will give you the information needed to get your problem handled correctly and efficiently.
- Have your Ascensia *ELITE*[™] Meter, Ascensia *ELITE*[™] Check Strip, and a box of Ascensia *ELITE*[™] Test Strips (including the Code Strip) nearby when you phone. It would also be helpful to have a bottle of Ascensia *ELITE*[™] Normal Control available.
- Fill out the **check list** on the following page before calling!

Check List

1. The Meter serial number is: _____
2. The date and time the problem occurred was: _____
3. I tested with both the Check Strip and with the Ascensia *ELITE*TM Normal control:
Yes _____ No _____
 - The Check Strip result was: _____
 - The Normal Control result was: _____
4. The Normal Control Range listed on the end flap of the Test Strip carton is: _____ to _____
5. The Function Number being displayed on the Meter is: _____
The Function Number on the Code Strip and on the Test Strip foil packet currently being used is: _____

6. Please include the lot number and expiration date of the Test Strips and Controls currently being used. For the Test Strips, this information is printed on the carton as well as on the back of each foil packet. For the Controls, this information is printed on the bottle.

	<u>Test Strips</u>	<u>Control</u>
Lot Number:	_____	_____
Expiration Date:	_____	_____

1. PRESERVATIVES:

- **Fluoride** – Do not use fluoride as a preservative for blood specimens.
- **Iodoacetic acid** – Do not use iodoacetic acid as a preservative for blood specimens.

2. **ARTERIAL SPECIMEN:** Arterial specimens with pO_2 of >200 mmHg had a small bias (average of -6%) on the Ascensia *ELITE*[™] in comparison to the laboratory method glucose results. The reference range for arterial pO_2 is 80 to 110 mmHg (11.04 to 14.36 kPa).
3. **NEONATAL USE:** This Ascensia *ELITE*[™] Blood Glucose Meter is not qualified for testing neonatal blood specimens.
4. **METABOLITES:** Reducing substances such as ascorbic acid and uric acid at normal blood concentrations do not significantly affect Ascensia *ELITE* Blood Glucose Test results.
5. **HEMATOCRIT:** At normal glucose levels, Ascensia *ELITE* Test Strip results are not significantly affected by hematocrits in the range of 20% to 60%. At glucose levels above 300 mg/dL (16.7 mmol/L), hematocrit levels above 55% will cause lowered results.
6. **LIPEMIC SPECIMEN:** Cholesterol up to 500 mg/dL (13.0 mmol/L) or triglycerides up to 3000 mg/dL (33.9 mmol/L) do not significantly affect the results. Glucose values however, in specimens beyond these levels should be interpreted with caution.

CONTRAINDICATION: Fingerstick blood glucose testing may not be clinically appropriate when peripheral blood flow is decreased. Shock, severe hypotension, hyperosmolar hyperglycemia and occurrence of severe dehydration are examples of clinical conditions which may adversely affect the measurement of glucose in peripheral blood.¹

¹Atkin S, Jaker MA, Chorost MI, Reddy S, Fingerstick Glucose Determination in Shock. *Annals of Internal Medicine* 1991, 114: 1020–24.

Test Sample: Whole blood

Test Result: Plasma/Serum glucose

Reaction Chamber

Sample Volume: 2 μ L

Measuring Range: 20–600 mg/dL (1.1–33.3 mmol/L)

Measuring Time: 30 Seconds

Calibration Curve: Selected by use of the Code Strip

Memory Feature: Stores last 20 test results

Temperature Compensation: Automatic compensation with built-in thermistor

Battery Type: 3-volt lithium batteries (DL or CR2032) – 2

Battery Life: Approximately 1000 Tests (1 year of average use)

Operating Temperature Range: 50°–104° F (10°–40° C)

Humidity: 20–80% RH

Dimensions: 81 x 51 x 14 mm (3¹¹/₃₂ x 2 x ⁹/₁₆ in.)

Weight: Approximately 50g (1.75 ounces)

Alarm: “Beeps” whenever a Check Strip, Code Strip or Test Strip is inserted into the Meter, when the Test Strip is completely filled with blood or Control Solution, and when a test result appears in the Display.

Specifications (continued)

TURNING OFF THE ALARM:

We **strongly advise** you to leave the alarm on, as it helps tell you when the Test Strip is adequately filled with blood or control solution. However, you can operate your Meter discreetly by turning off the alarm.

MATERIALS REQUIRED:

Ascensia *ELITE*™ Check Strip or Test Strip (both strips will work).





1 Insert Check or Test Strip fully into Meter.

2 Turn the Meter over and *open and shut* the battery door.

3 Then *quickly* turn the Meter back over (display is blank), *remove and reinsert* strip within 10 seconds. **on** is displayed.

Immediately pull the strip out within 2 seconds. **off** is displayed. The alarm is now off.

To turn the alarm back on, repeat this procedure (#1-2-3). The Display sequence is reversed, **off** appears first and then **on**.



When calling or writing for supplies be sure to include the number with the name of the replacement part or accessory item.

REPLACEMENT PARTS

Part Number*	Item
40030030	3-volt Lithium Battery (DL or CR2032) – 2 (Available at most Camera Stores)
99M39019	Ascensia <i>ELITE</i> ™ User Guide
50184221	Quick Reference Card

These parts are available in the United States from:

**Bayer Corporation
Customer Service Department
Order Entry P. O. Box 2004
Mishawaka, Indiana 46546**

*Part numbers are subject to change without notice

ACCESSORY ITEMS

Product Number	Item
3911A	Ascensia <i>ELITE</i> ™ Test Strips – 25's in foil
3918A	Ascensia <i>ELITE</i> ™ Test Strips – 50's in foil
3942A	Ascensia <i>ELITE</i> ™ Test Strips – 100's in foil
3940A	Ascensia <i>ELITE</i> ™ Low and High Controls
3972A	Ascensia <i>ELITE</i> ™ Normal Control
5509A	Ames Gluco System™ Lancets (100's)
5702M	GLUCOLET® Automatic Lancing Device
5700G	GLUCOLET® Endcaps – Regular
5701G	GLUCOLET® Endcaps – Super
6540B	MICROLET® Adjustable Lancing Device
3416	MICROLET Vaculance® Lancing Device
6546B	MICROLET® Lancets (100's)
6550B	MICROLET® Lancets (200's)

Accessory items may be obtained from Diabetes Care Centers, retail drug chains or other home healthcare distributors.

Limited Warranty

SECTION 7

Manufacturer's Warranty: Bayer Corporation ("Bayer") warrants to the original purchaser that this instrument will be free from defects in materials and workmanship for 5 years from the date of original purchase (except as noted below). During the stated 5-year period, Bayer shall, at no charge, replace a unit found to be defective with an equivalent or current version of the owner's model.

Limitations of Warranty: This warranty is subject to the following exceptions and limitations:

1. A 90-day warranty only will be extended for consumable parts and/or accessories.
2. This warranty is limited to replacement due to defects in parts or workmanship. Bayer shall not be required to replace any units which malfunction or are damaged due to abuse, accidents, alteration, misuse, neglect, maintenance by other than Bayer, or failure to operate the

instrument in accordance with instructions. Further, Bayer assumes no liability for malfunction or damage to Bayer instruments caused by the use of reagents other than reagents (i.e., Ascensia *ELITE*™ Test Strips) manufactured or recommended by Bayer.

3. Bayer reserves the right to make changes in design of this instrument without obligation to incorporate such changes into previously manufactured instruments.
4. **Important:** Bayer has no knowledge of the performance of the Ascensia *ELITE*™ Blood Glucose Meter when used with any Test Strips other than Ascensia *ELITE* Test Strips, and therefore makes no warranty of the performance of the Ascensia *ELITE* Meter when used with any Test Strips other than Ascensia *ELITE* Test Strips or when the Ascensia *ELITE* Test Strip is altered or modified in any manner.

Warranty (continued)

SECTION 7

BAYER MAKES NO OTHER EXPRESS WARRANTY FOR THIS PRODUCT. THE OPTION OF REPLACEMENT, DESCRIBED ABOVE, IS BAYER'S ONLY OBLIGATION UNDER THIS WARRANTY.

IN NO EVENT SHALL BAYER BE LIABLE FOR INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES, EVEN IF BAYER HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

Some states do not allow the exclusion or limitation of incidental or consequential damages, so the above limitation or exclusion may not apply to you. This warranty gives you specific legal rights and you may also have other rights which vary from state to state.

For warranty service: Purchaser must contact the Customer Service Department of Bayer Corporation, by calling toll free 1-800-348-8100, for assistance and/or instructions for obtaining service of this instrument.

Ascensia *ELITE*[™] is a trademark of Bayer Corp. CLINILOG[®], GLUCOLET[®] and MICROLET Vaculance[®] are registered trademarks of Bayer Corporation, Elkhart, Indiana.



Bayer Corporation
Mishawaka, IN 46544 USA
www.bayercaresdiabetes.com
www.ascensia.com